



Product Service

# CERTIFICATE

No. Z1 18 06 56257 190

**Holder of Certificate:** **Leica Microsystems CMS GmbH**  
 Ernst-Leitz-Straße 17-37  
 35578 Wetzlar  
 GERMANY

**Production Facility(ies):** 30821

**Certification Mark:**



**Product:** **Laboratory Equipment (Microscope)**

**Model(s):** **DM1000 LED**

**Parameters:**

Microscope:	
Rated voltage:	12V
Rated frequency:	DC
Rated input power:	18W
Protection class:	III
Lamp classification group:	Exempt group

**Tested according to:** EN 61010-1:2010  
 EN 61010-2-101:2017

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

**Test report no.:** 028-713135642-000

**Valid until:** 2023-06-24

**Date,** 2018-06-26

  
 (Ralph Fischer)

Page 1 of 1

